**Calibration and Reconstruction Performances of the KLOE Electromagnetic Calorimeter** 



# P.Gauzzi, Rome University and INFN for the KLOE collaboration

IX Int.Conf.on Calorimetry in Particle Physics LAPP Annecy – 9-14 October 2000

## Outline

- The KLOE experiment and the EMC design
- Reconstruction and calibration of energy
- Reconstruction and calibration of timing
- Performances on simple physics examples



#### THE KLOE Experiment @ DAFNE



**Be beam pipe** (0.5 mm thick) **Instrumented permanent magnet quadrupoles** (32 PMT's) QCAL

**Drift chamber** (4 m Ø × 3.3 m) 90% helium 10% isobutane 12582/52140 sense/total wires

**Electromagnetic calorimeter** Lead/scintillating fibers

**Superconducting coil**  $B = 0.6 \text{ T} (\int B \, dl = 2.2 \text{ T} \cdot \text{m})$ 

DAFNE: e<sup>+</sup>e<sup>-</sup> collider at **j** peak; **Ö**s = 1020 MeV

The **KLOE** goal is to measure  $\text{Re}(\mathbf{e'}/\mathbf{e})$  to ~  $\mathbf{O}$  (10<sup>-4</sup>)

#### **KLOE CALORIMETER REQUIREMENTS**

- Reconstruct K<sub>S</sub>,K<sub>L</sub> ® p<sup>0</sup>p<sup>0</sup>
   vertices with ~ 1cm resolution
- Discriminate K<sub>L</sub> ® p<sup>0</sup>p<sup>0</sup> from K<sub>L</sub> ® p<sup>0</sup>p<sup>0</sup>p<sup>0</sup>p<sup>0</sup>
- Fast for triggering and Bhabha background rejection
- Provide useful information for particle identification (*K*<sub>m</sub>) rejection)

- **s**(E)/E ~5% /ÖE(GeV)
- High efficiency  $20 < E_g < 300 \text{ MeV}$
- **s**(t) ~70 ps / **Ö***E* (GeV)
- s<sub>x,y,z</sub> ~1 cm for photon conversion point
- Hermeticity

### THE CALORIMETER STRUCTURE

# Fine sampling lead/scintillating fibers calorimeter

- Volume Ratio Fiber:Lead 50:50
- Energy sampling fraction 13 %
- $X_0 = 1.6 \text{ cm} \text{ } r = 5.3 \text{ g/cm}^3$







24 barrel modules: 4.3m length
60 cells (5 layers) – 4.4×4.4 cm<sup>2</sup> granul.
2 × 32 endcap modules 10/15/30 cells
4880 read-out channels

Fiber choice: Kuraray SCSF-81 and Pol.Hi.Tech 0046 PM choice: Mesh Hamamatsu R5946 1.5"

### **ENERGY CALIBRATION with COSMICS**



• Online filter selects 100 Hz of golden mips in the whole detector

• Peak of the ADC spectrum = response at calorimeter center (MIP) (~1 day data taking ® 1000 events/cell ® 1,2% stat. accuracy)

• The attenuation curve is measured for each channel and used in the reconstruction procedure (w(z) = A  $e^{-z/l_1} + (1-A) e^{-z/l_2}$ )

#### **ENERGY CALIBRATION with E.M. SHOWERS**



#### LINEARITY IN ENERGY RESPONSE AND ENERGY RESOLUTION



E<sub>v</sub>(MeV)





#### TIMING CALIBRATION WITH COSMIC RAYS ... continue

• In 1 hour of data taking we collect enough statistics to calibrate t<sup>0</sup>'s and v<sub>eff</sub> with the following accuracy:

- 30 , 40 ps for each t<sup>0</sup>
- 0.3 % on v<sub>eff</sub> (average value 16.7 cm/ns)



**for a MIP in a cell** (~38MeV) :  $s(t) \sim 340 \text{ ps}$ corresponding to:  $s(t) \sim 64 \text{ ps} / \ddot{o}E(GeV)$ 



#### TIME RESOLUTION

Comparing the difference of timing between **g**events at small and large angle we estimate that of the 147 **ps** of constant term:

- 50 ps mis-calibration
- 55 ps bunch spread
- 120 ps machine time spread



#### **TIMING PERFOMANCES: identification of K<sub>1</sub> interacting on EMC**

## Clean signature (late neutral cluster) is used for $K_S$ tag





**RECONSTRUCTION AND ALIGNEMENT OF NEUTRAL VERTEX** 

## **CONCLUSIONS AND OUTLOOK**

• The KLOE Calorimeter has been kept in operation as a whole apparatus for more than 2.5 years in good operating conditions (the number of dead channels always below 0.1 % and overall good detector stability)

- The calibration procedures of energy and timing are working.
- Energy resolution of 5.7 % / Ö(E/GeV) measured
- Timing resolution of 54 ps /Ö (E/GeV) Å 50 ps achieved
- Good calibration stability observed in 6 months of operation.
- Reconstruction of masses and neutral vertices satisfactory
- Work is in progress to:
- 1) complete the setting of the time scale

2) correct the residual response non-linearity and the response along cracks regions

3) make fully automatic the calibration procedures